

STATEMENT OF FACT

Report No. CAL/14-388 (V)
(Customer PO # GLOBE-0001363728)

This document serves to verify that the following equipment(s) has/have been tested to standards laid down by Martin-Decker/Totco.


Customer : *Transocean*
Rig / Barge : Discoverer Seven Seas
Report Date : 12 August 2014
Job No. : NOV-532993
Model & Description : WE3-18, Type 150 Weight Indicator
System Serial No. : 0912309
Load Cell Model & Description : NA
Load cell Serial No. : NA
Capacity : 1,800,000 LBS (12 Lines)
Room Temperature : 23°C
Date of Calibration : 12 August 2014
Date of Next Calibration : 11 August 2015

Note : The above indicator has been calibrated at quadrant check points with reference to Drg No WE3, Rev DA:

**450 PSI @ 90°
900 PSI @ 180°
1350 PSI @ 270°
1800 PSI @ 360°**

Load Reference: The described item was verified using Pressure Gauge S/N **86TG0043** with certificate no. **PP095340-381** traceable to National Standard maintained by NPL & SPRING. This calibration is performed in accordance with BS EN 837 : 1998, clause 10.2.

Verified by:



Abdullah Hamsos
Workshop Technician

Reviewed by:



Kelson Ho
Service Supervisor